

STUDY TITLE

Evaluation of Antimicrobial Activity of UV Illumination /
Hydroxyl Generator

Test Organisms:

Aspergillus niger (ATCC 16404)

PRODUCT IDENTITY

Odorox Mobile Disinfection Unit Hydroxyl Generator

AUTHOR

Scott R. Steinagel, B.S.
Manager, Microbiology Laboratory Operations

STUDY COMPLETION DATE

January 23, 2009

PERFORMING LABORATORY

ATS Labs
1285 Corporate Center Drive, Suite 110
Eagan, MN 55121

SPONSOR

Safety Performance Solutions, Inc.
3908 Kingston Drive
Bismarck, ND 58503

PROJECT NUMBER

A07206

STUDY REPORT

GENERAL STUDY INFORMATION

Study Title: Evaluation of Antimicrobial Activity of UV Illumination / Hydroxyl Generator
Project Number: A07206
TRF Number: SPS01120908.CUST.2

TEST SUBSTANCE IDENTITY

Test Substance Name: Odorox Mobile Disinfection Unit Hydroxyl Generator

STUDY DATES

Date Sample Received: September 30, 2008
Study Initiation Date: January 6, 2009
Experimental Start Date: January 12, 2009
Experimental End Date: January 19, 2009
Study Completion Date: January 23, 2009

Test Organism	ATCC #	Culture Medium	Subculture Plate Medium
<i>Aspergillus niger</i>	16404	Sabouraud Agar (Modified)	Sabouraud Dextrose Agar

The microorganism used in this study was obtained from the American Type Culture Collection (ATCC), Manassas, Virginia.

Test Exposure: 48 hours, 72 hours, and 96 hours
Exposure Temperature: Room temperature (25.0-31.6°C).
Number of Carriers Tested/lot: Duplicate carriers per exposure time utilizing two carrier types, 1" x 1" stainless steel and 1" x 1" cotton fabric
Soil Load Description: No organic soil load required
Neutralizing Subculture Medium: Lethen Broth with 0.07% Lecithin and 0.5% Tween 80

EXPERIMENTAL DESIGN

An incubator (approximately 35" x 26" x 76.5") was prepared for testing by turning off all applicable fans and heat sources allowing the incubator to equilibrate to room temperature. The Odorox Mobile Disinfection Unit Hydroxyl Generator was placed into the incubator; the unit was powered on and was allowed to run for 72 minutes prior to placing the carriers in the incubator. Duplicate test carriers, per carrier type, per exposure time point were inoculated with a dried film of test culture and were placed within the incubator. Fabric carriers were allowed to hang freely, while stainless steel carriers were exposed within Petri dishes with the dish lids fully ajar. Following a 48 hour, 72 hour and 96 hour exposure, the carriers were neutralized, mixed and assayed for survivors. Side by side fabric and stainless steel quantitation control carriers were inoculated and dried as in the test. A single control carrier was neutralized immediately after drying (time zero). Additionally, duplicate control carriers were exposed for 48 hour, 72 hour and 96 hours, as in the test, under ambient conditions. Appropriate purity, carrier sterility, and neutralizing subculture medium sterility controls were performed. Percent and log₁₀ reductions were determined for the test carriers as compared to the quantitation control carriers.

TABLE 1: CONTROL RESULTS

Type of Control		Results
		<i>Aspergillus niger</i> (ATCC 16404)
Purity Control		Pure
Neutralizing Subculture Medium Sterility Control		No Growth
Carrier Sterility Control	Stainless Steel	No Growth
	Cotton Fabric	No Growth

**TABLE 2: EVALUATION OF QUANTITATION CONTROL CARRIER DATA
(TIME ZERO)**

Test Organism	Carrier type	CFU/carrier	Log ₁₀
<i>Aspergillus niger</i> (ATCC 16404)	Stainless Steel	5.10 x 10 ⁵	5.708
	Cotton Fabric	4.94 x 10 ⁵	5.694

CFU = Colony Forming Unit

**TABLE 3: EVALUATION OF QUANTITATION CONTROL CARRIER DATA
(FOLLOWING EXPOSURE)**

Test Organism	Exposure Time	Carrier type	Average CFU/carrier	Average Log ₁₀
<i>Aspergillus niger</i> (ATCC 16404)	48 hours	Stainless Steel	5.4×10^4	4.73
		Cotton Fabric	1.3×10^5	5.12
	72 hours	Stainless Steel	6.5×10^4	4.81
		Cotton Fabric	1.9×10^5	5.27
	96 hours	Stainless Steel	6.3×10^4	4.80
		Cotton Fabric	2.15×10^5	5.332

CFU = Colony Forming Unit

TABLE 4: EVALUATION OF TEST CARRIER DATA

Test Substance	Test Organism	Exposure Time	Carrier type	Average CFU/carrier	Average Log ₁₀
Odorox Mobile Disinfection Unit Hydroxyl Generator	<i>Aspergillus niger</i> (ATCC 16404)	48 hours	Stainless Steel	2×10^2	2.2
			Cotton Fabric	1.5×10^4	4.17
		72 hours	Stainless Steel	2×10^1	1.3
			Cotton Fabric	7.6×10^3	3.88
		96 hours	Stainless Steel	$<2 \times 10^1$	<1.3
			Cotton Fabric	2.1×10^3	3.32

CFU = Colony Forming Unit

**TABLE 5: RELATIVE ORGANISM REDUCTION AS COMPARED TO SIDE BY SIDE
QUANTITATION CONTROL CARRIERS**

Test Substance	Test Organism	Exposure Time	Carrier type	Percent Reduction	Log ₁₀ Reduction
Odorox Mobile Disinfection Unit Hydroxyl Generator	<i>Aspergillus niger</i> (ATCC 16404)	48 hours	Stainless Steel	99.6%	2.5
			Cotton Fabric	88.5%	0.95
		72 hours	Stainless Steel	>99.9%	3.5
			Cotton Fabric	96.0%	1.39
		96 hours	Stainless Steel	>99.9%	>3.5
			Cotton Fabric	99.0%	2.01

**TABLE 6: OVERALL ORGANISM REDUCTION AS COMPARED TO TIME ZERO
QUANTITATION CONTROL CARRIERS**

Test Substance	Test Organism	Exposure Time	Carrier type	Percent Reduction	Log ₁₀ Reduction
Odorox Mobile Disinfection Unit Hydroxyl Generator	<i>Aspergillus niger</i> (ATCC 16404)	48 hours	Stainless Steel	>99.9%	3.5
			Cotton Fabric	97.0%	1.52
		72 hours	Stainless Steel	>99.99%	4.4
			Cotton Fabric	98.5%	1.81
		96 hours	Stainless Steel	>99.99%	>4.4
			Cotton Fabric	99.6%	2.37

ANALYSIS

Odorox Mobile Disinfection Unit Hydroxyl Generator, demonstrated a 99.6% (2.5 log₁₀) relative reduction on stainless steel and an 88.5% (0.95 log₁₀) relative reduction on cotton fabric for *Aspergillus niger* (ATCC 16404) following a 48 hour exposure period, greater than a 99.9% (3.5 log₁₀) relative reduction on stainless steel and a 96.0% (1.39 log₁₀) relative reduction on cotton fabric for *Aspergillus niger* (ATCC 16404) following a 72 hour exposure period, and greater than a 99.9% (>3.5 log₁₀) relative reduction on stainless steel and a 99.0% (2.01 log₁₀) relative reduction on cotton fabric for *Aspergillus niger* (ATCC 16404) following a 96 hour exposure period, as compared to side by side quantitation control carriers, when tested at room temperature (25.0-31.6°C).


Odorox Mobile Disinfection Unit Hydroxyl Generator, demonstrated greater than a 99.9% (3.5 log₁₀) overall reduction on stainless steel and a 97.0% (1.52 log₁₀) overall reduction on cotton fabric for *Aspergillus niger* (ATCC 16404) following a 48 hour exposure period, greater than a 99.99% (4.4 log₁₀) overall reduction on stainless steel and a 98.5% (1.81 log₁₀) overall reduction on cotton fabric for *Aspergillus niger* (ATCC 16404) following a 72 hour exposure period, and greater than a 99.99% (>4.4 log₁₀) overall reduction on stainless steel and a 99.6% (2.37 log₁₀) overall reduction on cotton fabric for *Aspergillus niger* (ATCC 16404) following a 96 hour exposure period as compared to the time zero quantitation control carriers, when tested at room temperature (25.0-31.6°C).

This study was performed following ATS Labs' Standard Operating Procedures (SOPs) and internal quality systems.

PROFESSIONAL PERSONNEL INVOLVED:

- | | |
|--------------------------|---|
| David Rottjakob, M.T. | - Director, Microbiology Services |
| Scott R. Steinagel, B.S. | - Manager, Microbiology Laboratory Operations |
| Joy Salverda, B.S. | - Research Scientist I |
| Adam W. Pitt, B.S. | - Research Assistant II |
| Matthew Sathe, B.S. | - Research Assistant II |
| Katherine C. Sager, B.S. | - Research Assistant I |
| Megan McDonald, B.S. | - Research Assistant I |
| John Kathrein | - Laboratory Assistant |

PREPARED BY:



 Scott R. Steinagel, B.S.
 Manager, Microbiology Laboratory Operations

1-23-09

 Date

REVIEWED BY:



 Judy Heidemann
 Quality Assurance Auditor

1-23-09

 Date

The use of the ATS Labs name, logo or any other representation of ATS Labs without the written approval of ATS Labs is prohibited. In addition, ATS Labs may not be referred to in any form of promotional materials, press releases, advertising or similar materials (whether by print, broadcast, communication or electronic means) without the express written permission of ATS Labs.